

WARNINGS and CAUTIONS for AR250 BEDSIDE HANDRAILS (APBRs)

ATTENTION!

The Consumers Product Safety Commission (CPSC), and the Food & Drug Administration (FDA), and the F3186-17 Standard have said that this Product (**Product**), and all other similar assistive bedside balance products are classified as Adult Portable Bed Rails (APBRs).

WARNING!

It is known that there is a risk of Entrapment when APBRs are used. Hartmobility Inc. (Hartmobility) is committed to providing our customers and users with information and education about what Entrapment is and how to avoid Entrapment when using our Products. To assist in understanding the Entrapment Risks, our leaflet called Entrapment Prevention Information is included in these instructions.

Entrapment risk occurs whenever a user of an APBR can have their head, neck, chest, or other body part trapped in the space between the bed, the mattress or any other object and the APBR. Entrapment can result in serious injury and death.

Before installing or using this Product it is important to consider carefully who will be using it. The F3186-17 ASTM Standard, the CPSC, and the FDA describe users who are particularly at risk of entrapment, suffocation and strangulation as; *“People with Alzheimer’s disease, dementia, or other neurological conditions, or those who are sedated, confused, or frail.”*

If the user has been diagnosed with any of these conditions, caregivers, and the installers, must understand and be aware of the Entrapment risks, and take full responsibility for the correct installation and use of this Product.

If you are not technically competent in installing this Product, we recommend that you hire a professional installer/contractor.

The following are examples show the text of the required ENTRAPMENT, STRANGULATION, SUFFOCATION AND FALL HAZARDS WARNING LABELS, as defined and required by the Consumers Product Safety Commission.

WARNING!

ENTRAPMENT, STRANGULATION, SUFFOCATION AND FALL HAZARDS

Gaps in and around this product can entrap and kill. People with Alzheimer’s disease or dementia, or those who are sedated, confused, or frail, are at increased risk of entrapment and strangulation. People attempting to climb over this product are at increased risk of injury or death from falls. Always make sure this product is properly secured to bed. If product can move away from bed or mattress, it can lead to entrapment and death.

WARNING!

SUFFOCATION/STRANGULATION/ENTRAPMENT HAZARD

If product is installed incorrectly or moves from its initial position gaps can occur which can entrap and kill. People with Alzheimer’s disease or dementia, or other neurological conditions, or those who are sedated, confused, or frail, are at increased risk of entrapment, suffocation and strangulation.

NEVER use unless product is tight against mattress without gaps and at least 12½” from headboard and footboard.

NEVER use with children.

NEVER use on toddler, bunk, water, or inflatable beds, or on beds with mattress toppers or soft compressible pads



ADDITIONAL WARNINGS & CAUTIONS

- a. Do not assemble, install or use this Hartmobility Inc. Product until you have read and understand these Warnings, our Entrapment Prevention Information document and the Product Assembly & Installation Instructions supplied. This will ensure that this Product is correctly assembled, installed and used.
- b. This Product is designed and intended to be used with Single to King sized beds that have an angle iron frame with a box spring, and a mattress combination.
- c. This Product can be used with innerspring or memory foam style mattresses from 6” to 16” thick.
- d. Do not use this Product with mattresses that are made of a material that will increase the potential for the mattress to sag on the edges and increase the gap between the top edge of the mattress and the Product. Examples of these mattresses are lightweight foam, air mattresses, any lightweight mattresses, any mattress that easily compress on the edges, and pillow tops on mattresses.
- e. Never use this Product without a bedframe.
- f. Do not use on articulating or adjustable beds.
- g. Do not install two of these Products on the same side of the bed.
- h. This Product must be installed securely to the angle iron bed frame and box spring.
- i. If the product is not properly secured, the mattress move into an unsafe position which increases the danger of entrapment. Follow the Assembly and Installation Instructions supplied for the correct way to install
- j. MAINTENANCE: It is important to inspect this Product frequently. Stop using immediately if this Product shifts out of its original installed position until it is readjusted into the correct position. See the Instructions for how to re-adjust correctly.
- k. Stop using immediately if this Product is damaged or broken, or if parts are missing.
- l. If this Product is to be used in a long-term care facility, a nursing home, an assisted living center, a supported independent living facility, or any similar institution, follow these instructions AND the institution’s APBR Bed Rails installation policies; their use of APBRs Bed Rail policies; and their rules and regulations. If this Product does not comply with the institution, do not install.
- m. If you have any questions about the assembly, installation or use of these Product(s); or any questions about the risks associated with these Product(s); or any other questions regarding any Hartmobility Inc. Product(s), please contact our Customer Support Help Desk at info@hartmobility.com
- n. In addition to contacting the manufacturer directly, consumers should report problems to the CPSC at its website: SaferProducts.gov or call 1-800-638-2772, or to the FDA at 1-800-332-1088.
- o. For further information, see www.fda.gov/bedsafety

WARNINGS and CAUTIONS for AR250 BEDSIDE HANDRAILS (APBRs)

Agreement Between Installer/User(s) and Hartmobility Inc.

Neither Hartmobility Inc., nor the Seller of this Product will be Liable for improper installation or use of this Product.

Warranty

For Warranty details, refer to the Warranty details on the Hartmobility Inc.

Website: www.hartmobility.com/Warranty or call 1-800-998-0774 ext. 3



Questions. Comments. Concerns.

Contact Customer Service today.

info@Hartmobility.com

www.SafetyGrip.com

www.Hartmobility.com

1(800) 998-0774

Hartmobility Inc
424 Williamson-Zebulon Rd.
Williamson, GA 30292

Hartmobility reserves the right to update and revise our products. Refer to our website for the most up-to-date version of this document.

Copyright © 2024 Hartmobility Inc. All Rights Reserved.

